

Evaluation of Technopath Controls on the ARCHITECT Family of Instruments

J. Litten¹, L. Nelson¹, K. Johnson², J. Shih² ¹Winchester Medical Center, Winchester, VA.; ²Abbott Laboratories, Abbott Park, IL, US

2015 AACC Annual Meeting, Atlanta, Georgia • 26 – 30 July, 2015

Abstract

Materials & Methods (cont'd)

laboratory testing to ensure that released results meet the required quality in regards to accuracy and precision of patient results. Consolidation of controls is a current trend in laboratories to simplify QC testing. Multiconstituent control panels (MCCs) offered by Technopath Manufacturing Ltd. cover a wide range of clinical chemistry two different levels. and immunoassay analytes for both serum and urine.

performance of the Multichem S Plus, Multichem IA Plus and Multichem U control panels on the ARCHITECT to the target value were evaluated.

Methods: The three control panels were evaluated for a minimum of 30 days. Testing was performed on two ARCHITECT c8000 and three ARCHITECT i2000_{SR} instruments. Data presented here are from the following serum clinical chemistry analytes: ALT, AST, total bilirubin, chloride, total cholesterol, creatinine (picrate), glucose, potassium, total protein, sodium, triglycerides and urea; the following immunoassay analytes: CEA, total PSA, free T3, free T4, TSH, troponin-I, total beta HCG, estradiol, ferritin, FSH, vitamin B12 and vitamin D; and the following clinical chemistry urine analytes: chloride, glucose, potassium, sodium and urea. All data were creatinine (picrate), glucose, potassium, sodium and urea. The Multichem S Plus and IA Plus panels are serum based with three control levels; the Multichem U panel is prepared from human urine with two control levels. All data were collected via AbbottLink, allowing for automated data retrieval. Means, standard deviations and ranges were calculated for all controls. Sigma Metrics were also calculated for each analyte.

with the Multichem S Plus control ranged from 0.46 to 5.33%. The %CV for the 6 clinical chemistry urine analytes with the Multichem U control ranged from 0.51 to 3.2%. For both control panels, the majority of the CVs Medical Council), and literature references². were less than 2%. The %CV for the 12 immunoassay analytes with the Multichem IA Plus control ranged from 1.34 to 18.87% (Tnl, Level 1); however the majority of the CVs were less than 5%. Overall, little variation was seen from instrument to instrument.

Introduction: Quality controls are an important part of 60 analytes. The IA Plus covers 45 analytes with three control levels. The U controls are provided in liquid form and stored at 2-8°C. Once the material is opened, it should be stored tightly capped at 2-8°C and is stable for 30 days unless otherwise stated in the lot specific data sheets. Thirteen analytes are included in the U control at

Testing was performed on two ARCHITECT c8000 and **Objective:** The goal of this study was to evaluate the three ARCHITECT i2000_{SR} instruments. The three control panels were evaluated for a minimum of 30 days in parallel with the lab's routine QC controls. The routine family of instruments. Precision and accuracy compared QC controls included Bio-Rad Unassayed Multiqual Chemistry controls, Liquichek Urine Chemistry, Liquichek Immunoassay Plus, Liquichek Fertility, and Liquichek Cardiac Plus LT.

> Data presented here are from the following serum clinical chemistry analytes: ALT, AST, total bilirubin, chloride, total cholesterol, creatinine (picrate), glucose, potassium, total protein, sodium, triglycerides and urea; the following immunoassay analytes: CEA, total PSA, free T3, free T4, TSH, troponin-I, total beta HCG, estradiol, ferritin, FSH, vitamin B12 and vitamin D; and the following clinical chemistry urine analytes: chloride, creatinine (picrate), collected via AbbottLink, allowing for automated data retrieval. Means, standard deviation and range were calculated for all controls.

Sigma-metrics were calculated for each analyte at each control level on each instrument using peer means and CV and Total Allowable Error (TEa) from the literature according to the following equation: Sigma-metric = (TEa - Bias)/CV. TEas were defined based on multiple Results: The %CV for the 12 clinical chemistry analytes sources, including the Clinical Laboratory Improvement Amendments (CLIA), The Royal College of Pathologists of Australasia (RCPA), Desirable Biological Variation Database¹, Rilibak (Guidelines of the German Federal

Results (cont'd)

Multichem S Plus

Assay	Units	Level	Inst	Ν	Expected Mean	Mean	SD	CV	
			<i>c</i> 8000#1	49	30.3	28	1.13	4.02	
ALT		Level 1	c8000#2	51	30.3	28.4	1.14	4.02	
	U/L		<i>c</i> 8000#1	49	112	108.8	1.72	1.58	
	0/L	Level 2	<i>c</i> 8000#2	53	112	108.4	1.51	1.39	
		Level 3	<i>c</i> 8000#1	47	240	231.3	4.71	2.04	
			<i>c</i> 8000#2	51	240	230.1	3.65	1.59	
AST	U/L	Level 1	c8000#1	52	43.9	43.4	0.86	1.97	
			c8000#2 c8000#1	51 51	43.9 128	42.3 127.4	0.75 1.58	1.78 1.24	
		Level 2	c8000#1	53	128	127.4	1.23	0.98	
			c8000#2	49	255	251.4	1.89	0.75	
		Level 3	c8000#2	51	255	247.5	1.72	0.7	
Chloride		1	c8000#1	55	85.1	84.77	0.65	0.77	
		Level 1	c8000#2	47	85.1	85.49	0.68	0.8	
	mmol/L	Level 2	<i>c</i> 8000#1	51	95.6	94.81	0.8	0.84	
omonae	1111101/ E	LOVOIZ	<i>c</i> 8000#2	48	95.6	95.23	0.76	0.79	
		Level 3	<i>c</i> 8000#1	49	110	109.09	0.77	0.7	
			c8000#2	47	110	109.23	0.58	0.53	
		Level 1	c8000#1	46	103	104.5	0.96	0.92	
			c8000#2	45	103	105	2.07	1.97	
Cholesterol	mg/dL	Level 2	c8000#1 c8000#2	46 46	157 157	161.1 160.5	1.83 2.45	1.14 1.52	
			<i>c</i> 8000#2 <i>c</i> 8000#1	46	248	253.7	3.08	1.52	
		Level 3	c8000#1	45	248	253.7	2.37	0.94	
			c8000#2	46	0.697	0.7	0.01	1.91	
		Level 1	c8000#2	46	0.697	0.7	0.02	2.41	
Creatinine			c8000#1	49	1.97	1.99	0.03	1.6	
(picrate)	mg/dL	Level 2	<i>c</i> 8000#2	48	1.97	2	0.04	1.94	
		Level 3	c8000#1	48	5.02	6.09	0.07	1.15	
		Level 3	<i>c</i> 8000#2	46	5.02	6.11	0.07	1.18	
	mg/dL	Level 1	<i>c</i> 8000#1	49	50.8	51.6	0.54	1.05	
Glucose		Leven	<i>c</i> 8000#2	47	50.8	51.8	0.58	1.12	
		Level 2	<i>c</i> 8000#1	50	133	133	1.27	0.95	
			c8000#2	48	133	134.7	1.3	0.97	
		Level 3	c8000#1	48	292	287.7	2.35	0.82	
		Level 1	c8000#2 c8000#1	47 55	292 2.42	292	3.16 0.03	1.08 1.18	
			c8000#1	47	2.42	2.57 2.57	0.03	0.98	
			c8000#2	51	3.77	3.94	0.020	0.74	
Potassium	mmol/L	Level 2	c8000#2	48	3.77	3.94	0.04	0.88	
			<i>c</i> 8000#1	50	6.76	6.87	0.06	0.86	
		Level 3	<i>c</i> 8000#2	47	6.76	6.88	0.03	0.46	
		Level 1	<i>c</i> 8000#1	55	121	120.7	1.13	0.93	
	mmol/L	Leven	c8000#2	47	121	121.26	0.96	0.79	
Sodium		Level 2	<i>c</i> 8000#1	51	141	140.56	1.11	0.79	
Couldin			<i>c</i> 8000#2	48	141	141.14	1.24	0.88	
		Level 3	c8000#1	50	161	161.18	1.2	0.74	
			c8000#2	47	161	161.68	0.88	0.55	
Total Bilirubin		Level 1	c8000#1 c8000#2	46 43	0.998	0.94	0.05	5.33 4.06	
	mg/dL		c8000#2	43	2.64	2.48	0.04	2.81	
		Level 2	c8000#1	48	2.64	2.40	0.07	3.13	
			c8000#1	46	5.55	4.84	0.19	3.89	
		Level 3	<i>c</i> 8000#2	47	5.55	4.88	0.19	3.98	
Total protein			<i>c</i> 8000#1	58	4.61	4.66	0.04	0.81	
	g/dL	Level 1	<i>c</i> 8000#2	53	4.61	4.68	0.04	0.83	
		Level 2	<i>c</i> 8000#1	57	6.52	6.61	0.06	0.83	
		LeverZ	<i>c</i> 8000#2	55	6.52	6.64	0.05	0.74	
		Level 3	c8000#1	55	8.75	8.8	0.04	0.5	
			c8000#2	53	8.75	8.85	0.04	0.47	
Triglycerides	mg/dL	Level 1	c8000#1	50 47	55.3	59.1	1.19	2.01	
			c8000#2 c8000#1	47 48	55.3 133	59.9 136.6	2.15 1.85	3.59 1.36	
		Level 2	c8000#1	40	133	138.5	2.05	1.30	
		Level 3	c8000#2	40	223	223.6	1.96	0.88	
			c8000#2	47	223	225.1	3.07	1.37	
			<i>c</i> 8000#1	43	8.54	8.3	0.28	3.33	
		Level 1	<i>c</i> 8000#2	48	8.54	8.3	0.18	2.18	
Urea	mg/dL	Louis LO	<i>c</i> 8000#1	44	39.2	38.6	0.88	2.29	
		Level 2	<i>c</i> 8000#2	50	39.2	38.8	0.78	2.02	
		Level 3	<i>c</i> 8000#1	42	63.8	62.8	1.44	2.28	
		200010	<i>c</i> 8000#2	48	63.8	63	1.15	1.82	

Multichem IA Plus

Assay	Units	Level	Inst	Ν	Expected Mean	Mean	SD	C۷
CEA		Level 1	<i>i</i> 2000#1	37	2.22	2.12	0.11	5.29
	ng/mL	Level 2	<i>i</i> 2000#1	37	16.8	16.58	0.67	4.01
		Level 3	<i>i</i> 2000#1	37	43.3	43.59	1.57	3.6
Estradiol		Level 1	<i>i</i> 2000#2	39	48.4	52	3.25	6.25
	pg/mL	Level 2	<i>i</i> 2000#2	37	159	165.4	3.6	2.18
		Level 3	<i>i</i> 2000#2	37	515	528.5	7.06	1.34
Ferritin	ng/mL		<i>i</i> 2000#1	40	20.3	21.03	0.95	4.5
		Level 1	<i>i</i> 2000#3	28	20.3	19.82	0.91	4.59
			<i>i</i> 2000#1	40	160	160.29	4.28	2.67
		Level 2 Level 3	<i>i</i> 2000#3	26	160	158.85	4.93	3.1
			<i>i</i> 2000#0	40	326	331.26	9.68	2.92
			<i>i</i> 2000#1	28	326	331.52	11.75	3.55
		Level 1	<i>i</i> 2000#3	35	2.09	1.7	0.09	5.47
Erec T2	n a /ml							
Free T3	pg/mL	Level 2	<i>i</i> 2000#1	35	4.33	4.2	0.16	3.81
		Level 3	<i>i</i> 2000#1	36	9	9.1	0.33	3.6
			<i>i</i> 2000#1	48	0.57	0.5	0.03	5.98
		Level 1	<i>i</i> 2000#2	39	0.57	0.53	0.02	3.81
			<i>i</i> 2000#3	47	0.57	0.53	0.03	5.07
			<i>i</i> 2000#1	47	1.69	1.68	0.04	2.36
Free T4	ng/dL	Level 2	<i>i</i> 2000#2	38	1.69	1.71	0.04	2.25
			<i>i</i> 2000#3	45	1.69	1.69	0.05	3.17
			<i>i</i> 2000#1	47	3.05	2.98	0.1	3.37
		Level 3	<i>i</i> 2000#2	38	3.05	3.05	0.1	3.28
			<i>i</i> 2000#3	45	3.05	2.94	0.15	4.93
	mIU/mL	Level 1	<i>i</i> 2000#2	38	6.2	6.1	0.18	3.02
FSH		Level 2	<i>i</i> 2000#2	37	17.8	17.4	0.6	3.46
		Level 3	<i>i</i> 2000#2	37	44.6	42.9	1.3	3.03
		Levero	<i>i</i> 2000#1	43	0.05	0.04	0	8.3
								11.2
Tnl		Level 1	<i>i</i> 2000#2	39	0.05	0.05	0.01	
	ng/mL		<i>i</i> 2000#3	34	0.05	0.05	0.01	18.8
		Level 2	<i>i</i> 2000#1	43	0.4	0.36	0.01	2.91
			<i>i</i> 2000#2	37	0.4	0.39	0.01	2.5
			<i>i</i> 2000#3	32	0.4	0.39	0.02	6.44
			<i>i</i> 2000#1	43	3.27	2.94	0.16	5.54
		Level 3	<i>i</i> 2000#2	37	3.27	2.99	0.07	2.36
			<i>i</i> 2000#3	34	3.27	3.035	0.17	5.46
		Level 1	<i>i</i> 2000#1	36	5.78	4.17	0.3	7.08
		Leven	<i>i</i> 2000#3	25	5.78	3.9	0.26	6.63
Total			<i>i</i> 2000#1	38	N/A*	23.34	1.24	5.31
beta-hCG	mIU/mL	Level 2	<i>i</i> 2000#3	27	N/A	24.28	1.32	5.46
			<i>i</i> 2000#1	38	505	437.02	12.4	2.84
		Level 3				422.21	10.51	
			<i>i</i> 2000#3	28	505			2.49
		Level 1			505 0.53			
Total PSA	na/ml	Level 1 Level 2	<i>i</i> 2000#2	34	0.53	0.47	0.02	4.01
Total PSA	ng/mL	Level 2	<i>i</i> 2000#2 <i>i</i> 2000#2	34 36	0.53 4.52	0.47 4.22	0.02 0.15	4.01 3.49
Total PSA	ng/mL		i2000#2 i2000#2 i2000#2	34 36 36	0.53 4.52 20.5	0.47 4.22 19.65	0.02 0.15 0.73	4.01 3.49 3.69
Total PSA	ng/mL	Level 2 Level 3	i2000#2 i2000#2 i2000#2 i2000#1	34 36 36 39	0.53 4.52 20.5 0.13	0.47 4.22 19.65 0.12	0.02 0.15 0.73 0.01	4.01 3.49 3.69 3.84
Total PSA	ng/mL	Level 2	i2000#2 i2000#2 i2000#1 i2000#1	34 36 36 39 36	0.53 4.52 20.5 0.13 0.13	0.47 4.22 19.65 0.12 0.13	0.02 0.15 0.73 0.01 0.01	4.01 3.49 3.69 3.84 4.01
Total PSA	ng/mL	Level 2 Level 3	i2000#2 i2000#2 i2000#1 i2000#1 i2000#2 i2000#3	34 36 39 36 36 34	0.53 4.52 20.5 0.13 0.13 0.13	0.47 4.22 19.65 0.12 0.13 0.12	0.02 0.15 0.73 0.01 0.01 0.01	4.01 3.49 3.69 3.84 4.01 4.09
		Level 2 Level 3 Level 1	i2000#2 i2000#2 i2000#1 i2000#2 i2000#3 i2000#1	34 36 39 36 34 40	0.53 4.52 20.5 0.13 0.13 0.13 4.09	0.47 4.22 19.65 0.12 0.13 0.12 3.92	0.02 0.15 0.73 0.01 0.01 0.01 0.11	4.01 3.49 3.69 3.84 4.01 4.09 2.82
Total PSA	ng/mL ulU/mL	Level 2 Level 3	i2000#2 i2000#2 i2000#1 i2000#2 i2000#3 i2000#1 i2000#2	34 36 39 36 34 34 40 38	0.53 4.52 20.5 0.13 0.13 0.13 4.09 4.09	0.47 4.22 19.65 0.12 0.13 0.12 3.92 3.95	0.02 0.15 0.73 0.01 0.01 0.01 0.11 0.09	4.01 3.49 3.69 3.84 4.01 4.09 2.82 2.16
		Level 2 Level 3 Level 1	i2000#2 i2000#2 i2000#1 i2000#3 i2000#3 i2000#1 i2000#2 i2000#3	34 36 39 36 34 40 38 32	0.53 4.52 20.5 0.13 0.13 0.13 4.09 4.09 4.09	0.47 4.22 19.65 0.12 0.13 0.12 3.92 3.95 3.99	0.02 0.15 0.73 0.01 0.01 0.01 0.11 0.09 0.15	4.01 3.49 3.69 3.84 4.01 4.09 2.82 2.16 3.79
		Level 2 Level 3 Level 1 Level 2	i2000#2 i2000#2 i2000#1 i2000#3 i2000#3 i2000#1 i2000#3 i2000#3	34 36 39 36 34 40 38 32 39	0.53 4.52 20.5 0.13 0.13 0.13 4.09 4.09 4.09 23.7	0.47 4.22 19.65 0.12 0.13 0.12 3.92 3.95 3.99 23.6	0.02 0.15 0.73 0.01 0.01 0.11 0.09 0.15 0.98	4.01 3.49 3.69 3.84 4.01 4.09 2.82 2.16 3.79 4.13
		Level 2 Level 3 Level 1	i2000#2 i2000#2 i2000#1 i2000#3 i2000#3 i2000#1 i2000#3 i2000#1 i2000#2	 34 36 39 36 34 40 38 32 39 37 	0.53 4.52 20.5 0.13 0.13 0.13 4.09 4.09 4.09 23.7 23.7	0.47 4.22 19.65 0.12 0.13 0.12 3.92 3.95 3.99 23.6 23.59	0.02 0.15 0.73 0.01 0.01 0.11 0.09 0.15 0.98 0.58	4.01 3.49 3.69 3.84 4.01 4.09 2.82 2.16 3.79 4.13 2.45
		Level 2 Level 3 Level 1 Level 2	i2000#2 i2000#2 i2000#1 i2000#3 i2000#3 i2000#1 i2000#3 i2000#1 i2000#2 i2000#3	 34 36 39 36 34 40 38 32 39 37 34 	0.53 4.52 20.5 0.13 0.13 0.13 4.09 4.09 4.09 23.7	0.47 4.22 19.65 0.12 0.13 0.12 3.92 3.95 3.99 23.6 23.59 23.85	0.02 0.15 0.73 0.01 0.01 0.11 0.09 0.15 0.98 0.58 0.83	4.01 3.49 3.69 3.84 4.01 4.09 2.82 2.16 3.79 4.13 2.45 3.46
		Level 2 Level 3 Level 1 Level 2 Level 3	i2000#2 i2000#2 i2000#1 i2000#3 i2000#3 i2000#1 i2000#3 i2000#1 i2000#2	 34 36 39 36 34 40 38 32 39 37 	0.53 4.52 20.5 0.13 0.13 0.13 4.09 4.09 4.09 23.7 23.7	0.47 4.22 19.65 0.12 0.13 0.12 3.92 3.95 3.99 23.6 23.59	0.02 0.15 0.73 0.01 0.01 0.11 0.09 0.15 0.98 0.58	4.01 3.49 3.69 3.84 4.01 4.09 2.82 2.16 3.79 4.13 2.45 3.46
		Level 2 Level 3 Level 1 Level 2	i2000#2 i2000#2 i2000#1 i2000#3 i2000#3 i2000#1 i2000#3 i2000#1 i2000#2 i2000#3	 34 36 39 36 34 40 38 32 39 37 34 	0.53 4.52 20.5 0.13 0.13 0.13 4.09 4.09 4.09 23.7 23.7 23.7	0.47 4.22 19.65 0.12 0.13 0.12 3.92 3.95 3.99 23.6 23.59 23.85	0.02 0.15 0.73 0.01 0.01 0.11 0.09 0.15 0.98 0.58 0.83	4.01 3.49 3.69 3.84 4.01 4.09 2.82 2.16 3.79 4.13 2.45 3.46 6.15
Total PSA TSH	ulU/mL	Level 2 Level 3 Level 1 Level 2 Level 3 Level 1	i2000#2 i2000#2 i2000#1 i2000#3 i2000#3 i2000#3 i2000#3 i2000#1 i2000#3 i2000#3	34 36 39 36 34 40 38 32 39 37 37 34 36	0.53 4.52 20.5 0.13 0.13 0.13 4.09 4.09 4.09 23.7 23.7 23.7 23.7 23.7	0.47 4.22 19.65 0.12 0.13 0.12 3.92 3.95 3.99 23.6 23.59 23.85 260.5	0.02 0.15 0.73 0.01 0.01 0.11 0.09 0.15 0.98 0.58 0.83 16.01	4.01 3.49 3.69 3.84 4.01 4.09 2.82 2.16 3.79 4.13 2.45 3.46 6.15 6.45
TSH		Level 2 Level 3 Level 1 Level 2 Level 3	i2000#2 i2000#2 i2000#1 i2000#3 i2000#3 i2000#1 i2000#3 i2000#1 i2000#3 i2000#1 i2000#1	 34 36 39 36 34 40 38 32 39 37 34 36 35 	0.53 4.52 20.5 0.13 0.13 0.13 4.09 4.09 4.09 23.7 23.7 23.7 23.7 23.7 274 274	0.47 4.22 19.65 0.12 0.13 0.12 3.92 3.95 3.99 23.6 23.59 23.85 260.5 271.3	0.02 0.15 0.73 0.01 0.01 0.11 0.09 0.15 0.98 0.58 0.83 16.01 17.51	4.01 3.49 3.69 3.84 4.01 4.09 2.82 2.16 3.79 4.13 2.45 3.46 6.15 6.45 5.42
TSH	ulU/mL	Level 2 Level 3 Level 1 Level 2 Level 3 Level 1 Level 1	i2000#2 i2000#2 i2000#1 i2000#3 i2000#3 i2000#1 i2000#3 i2000#1 i2000#3 i2000#1 i2000#1 i2000#2	 34 36 39 36 34 40 38 32 39 37 34 36 35 35 	0.53 4.52 20.5 0.13 0.13 0.13 4.09 4.09 4.09 23.7 23.7 23.7 23.7 274 274 274 547 547	0.47 4.22 19.65 0.12 0.13 0.12 3.92 3.95 3.99 23.6 23.59 23.85 23.59 23.85 260.5 271.3 544.8	0.02 0.15 0.73 0.01 0.01 0.11 0.09 0.15 0.98 0.58 0.83 16.01 17.51 29.55 23.87	4.01 3.49 3.69 3.84 4.01 4.09 2.82 2.16 3.79 4.13 2.45 3.46 6.15 6.45 5.42 4.16
TSH	ulU/mL	Level 2 Level 3 Level 1 Level 2 Level 3 Level 1	 i2000#2 i2000#2 i2000#1 i2000#3 i2000#3 i2000#3 i2000#3 i2000#3 i2000#3 i2000#1 	34 36 39 36 34 40 38 32 39 37 34 36 35 35 34 34	0.53 4.52 20.5 0.13 0.13 0.13 4.09 4.09 4.09 23.7 23.7 23.7 23.7 23.7 23.7 23.7 23.7	0.47 4.22 19.65 0.12 0.13 0.12 3.92 3.95 3.99 23.6 23.59 23.85 23.85 260.5 271.3 544.8 574.4 959.3	0.02 0.15 0.73 0.01 0.01 0.11 0.09 0.15 0.98 0.58 0.83 16.01 17.51 29.55 23.87 71.2	4.01 3.49 3.69 3.84 4.01 4.09 2.82 2.16 3.79 4.13 2.45 3.46 6.15 6.45 5.42 4.16 7.42
TSH	ulU/mL	Level 2 Level 3 Level 1 Level 2 Level 3 Level 1 Level 2 Level 2	 i2000#2 i2000#2 i2000#1 i2000#3 i2000#3 i2000#3 i2000#3 i2000#3 i2000#3 i2000#1 	34 36 39 36 34 40 38 32 39 37 34 35 35 35 35 34 34 33	0.53 4.52 20.5 0.13 0.13 4.09 4.09 4.09 23.7 23.7 23.7 23.7 23.7 23.7 23.7 23.7	0.47 4.22 19.65 0.12 0.13 0.12 3.92 3.95 23.95 23.65 23.59 23.85 260.5 271.3 544.8 574.4 959.3 1013.5	0.02 0.15 0.73 0.01 0.01 0.11 0.09 0.15 0.98 0.58 0.83 16.01 17.51 29.55 23.87 71.2 40.94	2.49 4.01 3.49 3.69 3.84 4.01 4.09 2.82 2.16 3.79 4.13 2.45 3.46 6.15 6.45 5.42 4.16 7.42 4.04 1.7
TSH	ulU/mL	Level 2 Level 3 Level 1 Level 2 Level 3 Level 1 Level 1	 i2000#2 i2000#2 i2000#1 i2000#3 i2000#3 i2000#3 i2000#3 i2000#3 i2000#3 i2000#1 	34 36 39 36 34 40 38 32 39 37 34 36 35 35 34 34	0.53 4.52 20.5 0.13 0.13 0.13 4.09 4.09 4.09 23.7 23.7 23.7 23.7 23.7 23.7 23.7 23.7	0.47 4.22 19.65 0.12 0.13 0.12 3.92 3.95 3.99 23.6 23.59 23.85 23.85 260.5 271.3 544.8 574.4 959.3	0.02 0.15 0.73 0.01 0.01 0.11 0.09 0.15 0.98 0.58 0.83 16.01 17.51 29.55 23.87 71.2	4.01 3.49 3.69 3.84 4.01 4.09 2.82 2.16 3.79 4.13 2.45 3.46 6.15 6.45 5.42 4.16 7.42

Conclusions: The Technopath S Plus, IA Plus and U controls performed well and demonstrated similar performance to the routine internal laboratory quality controls. The use of these MCCs reduced the number of controls required for the analytical quality control testing of both clinical chemistry and immunoassay analytes with no compromise to quality.

Background

Efficient analytical and quality control is important for testing laboratories to ensure that released results meet the required quality in regards to accuracy and precision. Consolidation of quality controls simplifies workflow in the laboratory. However, changing of quality controls necessitates careful evaluation by the laboratory. Technopath Manufacturing Ltd (Ballina, Ireland) recently introduced multi-constituent controls (MCC) for use with the Abbott ARCHITECT instruments that allows one control to be used for the majority of the assays available in the routine lab on the ARCHITECT instruments. The performance of the Technopath Multichem S Plus, Multichem IA Plus and Multichem U control panels were evaluated in comparison to the laboratory's current QC methods on two ARCHITECT c8000 and three ARCHITECT i2000_{SR} instruments at Winchester Medical Center in Winchester, VA.

Materials & Methods

Technopath Multichem S Plus, U and IA Plus controls are prepared from human serum or urine to which purified biochemical material (extracts of human and animal origin), chemicals, drugs, preservatives and stabilizers are added. The S Plus and IA Plus controls are provided in liquid form and stored frozen (-20 to -80°C)until use. Once thawed, the material is stored at 2-8°C and stable for 10 days unless otherwise stated in the lot specific data sheets. The S Plus controls have three levels covering

Results

The %CV for the 12 clinical chemistry analytes with the Multichem S Plus control ranged from 0.46% (Potassium) to 5.33% (Total Bilirubin).

The %CV for the 6 clinical chemistry urine analytes with the Multichem U control ranged from 0.51% (Chloride) to 3.2% (Urea). For both control panels, the majority of the CVs were less than 2%.

The %CV for the 12 immunoassay analytes with the Multichem IA Plus control ranged from 1.34% (Estradiol) to 18.87% (Troponin I. Level 1). The majority of the CVs were less than 5%. Overall, little variation was seen from instrument to instrument.

World class Six Sigma performance was seen with both Technopath and BioRad controls on the majority of analytes and control levels evaluated.

Analytical performance of the Technopath controls are summarized in the following tables.

Conclusions

- The Technopath S Plus, IA Plus and U controls performed well and demonstrated similar performance to the routine internal laboratory quality controls.
- The use of these MCCs reduced the number of controls required for the analytical quality control testing of both clinical chemistry and immunoassay analytes with no compromise to quality.

References

- 1. Ricos C, Alvarez V, Cava F, Garcia-Lario JV, Hernandez A, Jimenez CV, Minchinela J, Perich C, Simon M. "Current databases on biologic variation: pros, cons and progress." Scand J Clin Lab Invest 1999;59:491-500. 2014 update located at http://www.westgard.com/ biodatabase1.htm Last accessed 5/28/15.
- 2. Viljorn A, Singh DK, Farrinton K, Twomey PJ, Analytical quality goals for 25-vitamin D based on biological variation, J Clin Lab Analysis 2011 25:2:130-133.

ABBOTT ARCHITECT c8000 and i2000sR are trademarks of Abbott Laboratories.

Multichem U

Assay	Units	Level	Inst	Ν	Expected Mean	Mean	SD	%CV
Chloride	mmol/L	Level 1	c8000#1	44	87.2	87.9	0.59	0.68
			c8000#2	43	87.2	87.9	0.45	0.51
		Level 2	c8000#1	43	193	190.4	1.26	0.66
			c8000#2	42	193	189.1	1.61	0.85
Creatinine (picrate)	mg/dL	Level 1	c8000#1	42	66.1	63.64	1.53	2.4
			c8000#2	41	66.1	63.61	1.32	2.08
		Level 2	c8000#1	41	135	126.34	2.38	1.88
			c8000#2	40	135	125.98	2.49	1.98
	mg/dL	Level 1	c8000#1	43	35.9	36.5	1.08	2.95
Glucose			c8000#2	41	35.9	36.4	1	2.74
		Level 2	c8000#1	41	364	359.7	8.68	2.41
			c8000#2	40	364	360	9.59	2.66
	mmol/L	Level 1	c8000#1	44	16.2	16.2	0.13	0.83
Potassium			c8000#2	43	16.2	16.1	0.1	0.62
		Level 2	c8000#1	43	58.8	58.9	0.89	1.5
			c8000#2	42	58.8	58.8	0.43	0.74
	mmol/L	Level 1	c8000#1	44	81.8	80.5	0.76	0.95
Sodium			c8000#2	43	81.8	80.5	0.74	0.91
		Level 2	c8000#1	43	158	156.6	1.27	0.81
			c8000#2	42	158	157.7	1.12	0.71
Lives	mg/dL	Level 1	c8000#1	37	429	420.9	11.35	2.7
			c8000#2	43	429	417.4	13.37	3.2
Urea			c8000#1	36	1003	983.6	23.53	2.39
		Level 2	c8000#2	42	1003	977.6	27.33	2.8

Sigma Metrics Comparison

Sigma metrics were calculated for each analyte at each control level on each instrument for the BioRad and Technopath controls. The percent of controls performing at each sigma value are plotted below.

