



TECHNOPATH
CLINICAL DIAGNOSTICS

**SEROLOGY
CONTROLS**

Multichem[®] ID
INFECTIOUS DISEASE QC

Multichem[®] ID-COVID19 G

Third-Party Quality Control for IgG and IgM*
Antibodies to SARS-CoV-2



Features

- Ready to use liquid stable controls
- Technical and scientific support

Specifications

- 12 month shelf life at 2°C to 8°C.
- 21 day open vial stability at 2°C to 8°C.

Product	Description	Platform	Config.	Part Code
Multichem ID-COVID19 G	Positive Control	Roche Cobas	4 x 4mL	CVG300P
Multichem ID-COVID19 G	Positive Control	Abbott i Systems	4 x 4mL	CVG200P
Multichem ID-COVID19 G	Positive Control	Abbott i Systems, Roche Cobas, Ortho VITROS	4 x 2mL	CVG100P

Analytes

SARS-CoV-2 IgG
SARS-CoV-2 IgM*

For further information see Instructions For Use.

*IgM Lot Specific Customer Communication letter available for CVG200P. For further information, please contact: multichemcovid@technopathcd.com

THE QUALITY CONTROL COMPANY

TECHNOPATH CLINICAL DIAGNOSTICS
www.technopathcd.com
multichemcovid@technopathcd.com

MM_IDCOVID19_G_Digi_R3